## Heat Transfer Workshop 1 Introduction to Heat Flux Sensors

## Name \_\_\_\_

This series of workshops is designed to help you understand heat transfer and temperature. Temperature is what most people measure, but heat transfer is equally important for engineering applications. The workshops use a combination of thermocouples and heat flux sensors to measure both of these quantities simultaneously. The sensor output is read by a data acquisition board (DAQ) that is connected through a USB port to a laptop computer. A program is loaded on the computer to graphically display results and provide a control panel.

Heat transfer (q) is the movement of thermal energy as a function of time. Typical units are watts, the same as power, which is the movement of mechanical energy as a function time. Heat flux is heat transfer per surface area (q''), with units of watts per square meter. This is analogous to a power density. It is measured across a thin, flat sensor that is encapsulated in a plastic cover. It outputs a voltage (E), which is directly proportional to the heat flux according to a supplied calibration (S), typically given as microvolts per watt per square meter, q'' = E/S. The computer program automatically does this conversion when supplied with the sensor's calibration value.

To measure temperature two thermocouples are provided. One is attached as part of the heat flux sensor. The other is a separate set of wires, welded together at one end to form the thermocouple junction. Thermocouples measure the difference in temperature from this junction to the connection point of the other end of the wires. The DAQ automatically measures the connection temperature and adjusts the voltage output to give absolute temperature values.

Together, this allows direct measurement of both individual temperatures and temperature difference. Heat transfer is driven by temperature difference. There is always a source of thermal energy at a higher temperature that moves to an energy sink at a lower temperature. Consequently, it is important to know the source and sink temperatures and the resulting heat flux in an engineering system.

The heat flux sensor kit provides the components as shown in Fig. 1. Included is a heat flux sensor with a thermocouple, a second separate thermocouple, a DAQ, an aluminum coupon, an aluminum fin, a small piece of wood, a thin heater and a small piece of cloth. The



Fig. 1 Heat Flux Sensor Kit Components

heater is designed to take power from the DAQ to provide a source of heat for some of the workshops. Many of the workshops use the human body as a heat source, which allows you to feel the thermal process while measuring the results.

To make the measurements put the heat flux sensor between a heat source and a heat sink. This can be as simple as a desk and your hand or your arm and the surrounding air. The heat flux sensor can be taped to a surface or placed between two objects to hold it in place. Single-sided can be used over top of the sensor or double-sided tape can be used between the sensor and a surface. In either case the goal is to create good contact between the entire sensor and the surface to obtain a good reading.

## Heat Transfer Workshop 1 Introduction Assignment

## Name \_\_\_\_\_

1. Watch the introductory video on Heat and Temperature. https://youtu.be/1SX8u3qXo0A

2. Assemble your system. Make sure the wiring is correct as shown in Fig. 2. The placement of the colored wires is important. Starting from the left, the first two wires are the heat flux sensor (white then red). The next two wires are the thermocouple (blue then red). The next two are for the second thermocouple (blue then red). The last two wires are black in either order for the heater.

3. Make sure the proper computer programs are installed on your computer.

4. Plug in the DAQ to a USB port (at least 2.0 or higher).

5. Start the program. A screen should appear as shown in Fig. 3.

6. Toggle the switches on the screen to make sure that both the data acquisition and the heater are working. The red buttons should turn green when they are operating.

7. Record the heat flux and sensor temperature from a source to a sink. Note that the output from the heat flux sensor is directional. In one direction the heat flux should read positive and if the sensor is flipped over the output should be negative. The thermocouple, however, only measures a temperature at a location. It does not have positive and negative values. In all cases, if a wire is broken or disconnected, the output will be very large and unresponsive. Use the save button to create an excel file. This file will have four columns beginning with time in seconds, followed by the heat flux in  $W/m^2$ , the third column is the sensor temperature (°C), and the fourth is the second temperature (°C).



Fig. 2 Picture of DAQ wiring.

Measurement Description:



Fig. 3 Computer operating screen.

q" = \_\_\_\_\_ T = \_\_\_\_\_

Are these values what you expected (including sign and direction)? Explain why.